



**Glenammer**

**Glenammer Sieves Limited**

62 Viewfield Road, Ayr, KA8 8HH, UK

Tel: ++ 44 (0)1292 261444

Email: [sales@glenammer.com](mailto:sales@glenammer.com)

Web: [www.glenammer.com](http://www.glenammer.com)

## **Manufacturers Bostwick Consistometer Test and Mechanical\* Calibration Certificate**

This certifies that the Bostwick Consistometer has been fully tested and calibrated in accordance with Glenammer quality and test procedures on the date recorded below.

We recommend all units are returned annually for a service and calibration to prolong the life and quality of your product. Please contact us to arrange this.

Serial number: 1960

<b>Mechanical parts – functional</b>	<b>Pass v</b>
<b>Gate - operational</b>	<b>Pass v</b>
<b>Consistometer is and can be levelled</b>	<b>Pass v</b>
<b>Scale – clear and accurate</b>	<b>Pass v</b>
<b>Mechanical dimensions correct</b>	<b>Pass v</b>

Tested and calibrated by: John Anthony

Date: 25/10/2021

\*A mechanical calibration is carried out using a certified jig. This method ensures the bubble level is able to give an accurate reading when the unit is levelled using the adjustment screws alone.

Glenammer Sieves Ltd., Unit 3, 62 Viewfield Road, Ayr, Ayrshire, Scotland. KA8 8HH

Telephone: + 44 (0)1292-261444 Email: [sales@glenammer.com](mailto:sales@glenammer.com)